



www.magnet-physik.de Page 1/1



## STANDARD FIELD COIL NFH 63,4



## Standard Field Coil according to Helmholtz NFH 63,4

A standard field coil according to Helmholtz is used to generate a homogeneous static magnetic field. For example it is suitable to check or to calibrate sensors.

It consists of two compact partial coils that are split by a thin wall. Higher field strengths as with conventional Helmholtz coils can be reached because of the relatively large windings. Up to 80 kA/m (1 kOe) can be generated for short time periods. The homogeneous field region can be accessed in axial and radial direction.

A proprietary calibration certificate, which documents traceability of calibration to national standards, is issued free of charge on purchase. A calibration can alternately be carried out in our DKD accredited calibration laboratory. Periodic recalibration is recommended and can of course also be performed by our laboratory.

## Technical Data

Field strength constant  $k_H$ : approx. 20 kA/m per A (250 Oe per A) Flux density constant  $k_B$ : approx. 25 mT per A (250 G per A)

Maximum current: 4 A

Operation time: 25 %, max. 5 minutes at full current max. 12 mm diameter, max. 20 mm length

Openings: 50 mm axial diameter, 8 mm diameter radial hole on rear side

Resistance: approx.  $2 \cdot 10 \Omega$ 

Due to continuous product improvements specifications are subject to change without notice.